

## Certificate of Analysis

Product: ISE Detergent  
 Product Number: SM-B01417452D  
 Shelf Life: 3 Years  
 Storage Conditions: 2 - 25°C  
 Quantity: 2 x 100 mL

Parameter	Specified Range	Units
Hypochlorite	5.5 - 6.5	%
Potassium	0.85 - 0.95	mmol/L
Chloride	105 - 115	mmol/L
pH	11.5 - 12.5	pH units
Conductivity	75 - 85	mS/cm

Details
<b>Function:</b> For use on an Advia Analyzer to rinse the pipettes.
<b>Quality Specifications:</b> Passed
<b>Reagent Contents:</b> 6% Sodium Hypochlorite. <u>Recommended storage is refrigerated.</u>  For in vitro diagnostic use only.

Traceability		
Analyte	Standard	Test Method
Hypochlorite	OEM Reagent	Thiosulfate Titration
Potassium	NIST 918b	Flame Photometry
Chloride	NIST 919b	Titrimetric
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.