

Certificate of Analysis

Product: Standard B
Product Number: DD-61013
Shelf Life: 2.5 years
Storage Conditions: 18 – 25 ° C
Quantity: 100 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	100.0 ± 2	mmol/L
Potassium	1.80 ± 0.05	mmol/L
Chloride	72 ± 2	mmol/L

Details
Function: For the calibration of Na ⁺ , K ⁺ , and Cl ⁻ .
Quality Specifications: Passed
Reagent Contents: A 100 mL aqueous solution containing salts and preservative. For in vitro diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.