

Certificate of Analysis

Product: ISE Internal Reference
 Product Number: RO-4522320D
 Shelf Life: 3 Years
 Storage Conditions: 18 – 25 ° C
 Quantity: 5 X 600 mL


| Parameter | Specified Range | Units |
|------------|---|----------|
| Appearance | Clear, free of visible particulate matter | |
| pH | 7.40 - 7.52 | pH units |

| Details |
|---|
| Function: Internal standard for Na, K, Cl measurements on the ISE module. For use on Cobas c501 Analyzers. |
| Quality Specifications: Passed |
| Reagent Contents: An aqueous solution containing Triethanolamine and preservative. For in vitro diagnostic use only. |

Traceability

| Analyte | Standard | Test Method |
|---------|--------------------------------------|-------------|
| pH | Secondary Standard Traceable to NIST | pH Meter |

Certified by:



Kathy Fisher
 Director, Quality Assurance
 Diamond Diagnostics Inc.