

## Certificate of Analysis

Product: ISE Diluent  
 Product Number: RO-4522630D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25 ° C  
 Quantity: 5 X 300 mL


Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
pH	<b>7.35 - 7.65</b>	pH units

Details
Function: To dilute samples for Na, K, Cl measurements on the ISE module. For use on the Cobas c501 Analyzers.
Quality Specifications: Passed
Reagent Contents: An aqueous solution containing Boric acid and preservative. For in vitro diagnostic use only.

### Traceability

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics, Inc.